

(राष्ट्रीय पशु जैव प्रोद्योगिकी संस्थान)
National Institute of Animal Biotechnology

Corrigendum -- Change of Specifications

Please refer NIAB Tender Details as follows.

Tender ID : 2018_DBTEC_411989_1

Tender Reference Number : NIAB/SP/2018-19/38

Tender Title : Glove Box

The following changes may please be noted before submission of bids with respect to the tender details mentioned above.

Revised /New changes in specifications

Specifcaton in place of Existing old specifications --- Read As : - Annexure -1 (as attached below)

The specification mentioned below should be treated as revised specification and bid must be submitted accordingly .

Rest of the tender conditions remains same.

Manager (S&P)

NIAB-Hyderabad

Date:- 18/12/2018

Changes in Specification

Annexure-1

Specification No.	Existing Specification	Modified Specification
Point No 1	Specifications for Inert Gas Work Station (Glove box) <ul style="list-style-type: none">Inner Box Dimensions 900-950mm x 1200-1250mm x 775-800mm [H x L x D]: 2 glove ports, 220mm diameter	Specifications for Inert Gas Work Station (Glove box) <ul style="list-style-type: none">Inner Box Dimensions 900-950mm x 1200-1250mm x 775-800mm [H x L x D]: or all may be variable. 2 glove ports, 220mm diameter, or all may be variable.
Other Mandatory	<ul style="list-style-type: none">One Year Warranty.Glove box and Purifier should be from single manufacturerSupplier should have minimum of 10 Installation experience in India as per same configuration.	<ul style="list-style-type: none">One Year Warranty.Glove box and Purifier should be from single manufacturer