

(राष्ट्रीय पशु जैव प्रोद्योगिकी संस्थान)  
National Institute of Animal Biotechnology

**Corrigendum -- Change of Opening Date & Specifications**

*Please refer NIAB Tender Details as follows.*

Tender ID : 2020\_DBTEC\_581018\_1

Tender Reference Number : NIAB/SP/2020-21/22

Tender Title : Pure and Ultra-pure water Purification System

The following changes may please be noted before submission of bids with respect to the tender details mentioned above.

In place of old dates mentioned in Tender , please consider following dates.

Document Download End Date :- in place of Existing old date --- Read As :- 07/10/2020

Bid Submission End date : in place of Existing old date --- Read As :- 07/10/2020

Bid Opening Date in place of Existing old date --- Read As :- 08/10/2020

**Revised /New changes in specifications**

Specification in place of Existing old specifications for S No. 02, 06 ,09 and 13 --- **Read As** : -  
Annexure -1 (as attached below) The specification mentioned below should be treated as revised  
specification and bid must be submitted accordingly with revised quantity .

Rest of the tender conditions remains same.

Manager (S&P)  
NIAB-Hyderabad  
Date:- 14/09/2020

Point	Old specifications	Revised specifications
2 STAGE-2	Inbuilt Pre-treatment cartridge with anti-scaling compound, Activated carbon and 0.5u filter, Also should contain RFID tag for traceability.	Inbuilt Pre-treatment cartridge with anti-scaling compound, Activated carbon and 0.5u filter, Also should contain
6 STAGE -5	Type II water should pass through feed water specific cartridge for removal of trace contaminants and also should possess RFID tag	Type II water should pass through feed water specific cartridge for removal of trace contaminants
6 STAGE-5	Application Specific cartridges to remove ionic and organic contaminants to trace levels and also should possess RFID tag.	Application Specific cartridges to remove ionic and organic contaminants to trace levels
9	Flow Rate (L/hr).....10 l/h	Flow Rate (L/hr).....15-20 l/h
13	Valid CE Certificate mentioning the Model quoted must be submitted	Valid CE Certificate mentioning the Model quoted must be submitted or Water quality should be maintained the standards including: ASTM Type 1, CLSI and ISO 3696 Type 1.

Rest of the Specifications remains same.